

MADE IN EUROPE




This product has been developed and manufactured using both raw materials and with production in Europe.

Choose quality, save money, and protect the environment.

(UK) User Manual

01_20231222_UK

Master Glide Trauma Board



-
-

Article no.	HMI no.	Name - Size	UDI-DI/GTIN
MC 004-1030	103330	Master Glide Trauma Board	5744001972656
MC 004-1032	-	Master Glide Traume Board - 160	5744001975008

Regulation MDR 2017/745 for medical devices class 1



ISO 12 31 03 18

The Original Nordic Human Handling Concept



Developed & Manufactured by:



Thank you for choosing a product from master care A/S.

We recommend reading the user manual thoroughly before using the product (-s).

The information in this user manual applies to the shown products provided that the products have not been modified or customized and that the products have been used as recommended in this user manual.

If the customer changes, repairs or uses the product in other combinations than recommended by master care A/S the CE-approval no longer applies and the guarantee becomes void.

In case of doubt, contact master care A/S.

MDR (EU) 2017/745:

Products in this user manual is CE-approved and follows the harmonized standards. This means that the products are in compliance with the essential health and safety directives. The product(-s) comply with the standards in the Medical Device Regulation MDR (EU) 2017/745 – Class 1.

The intended use of the product/-s:

The product can be used for lying transfers of users between two surfaces. It is recommended to use sliding material and/or a draw sheet.

Proper use of the product:

Master Glide Trauma Board can be used when transferring between narrow beds e.g., X-ray beds or gurneys. Place the board halfway under the user after which the user slides over to the receiving bed. Nylon should be used if the transfer is considered difficult/hard or for users who are at risk of sustaining shear injuries.

Warning:

The board should not be used over longer unsupported distances than that indicated on the board.

Do not use for bariatric patients due to the board's narrow width.

Risk assessment:

Always make a risk assessment and a functional assessment of the user prior to using the product. This way using the product for transferring or moving a user, is ergonomic and legally sound.

Knowledge of the product/transferring:

It is important that the user of the product has been properly instructed in the use of the product and has read the user manual thoroughly prior to using the product.

master care A/S provide education in the use of products and on our website all user manuals can be found.

Life span:

The expected life span of the products with daily wear and tear is averagely 2-5 years. The products life span depends on the use and user's load.

Make visual inspections frequently and preferably before every use. Products with insufficient friction reduction, broken material or other deficiencies must be discarded. This way you will always keep a sound and healthy work environment for both helper and user.

Guarantee:

master care A/S provide a 2-year guarantee covering material defects and manufacturing defects if the product has been used correctly.

Unintended events and injuries when misusing the products:

Unintended events must always be reported to the correct national authority and to master care A/S or the distributor.

master care A/S takes no responsibility for personal injuries or product damages if the product(-s) have been used in contravention of the by the manufacturer recommended intended purpose or proper use of the product.



User Manual - Master Glide Trauma Board

How to use Master Glide Trauma Board:



1. Place the user along the edge of the mattress.



2. Tip the user slightly towards side lying position and place Master Glide Trauma under the head and pelvis.



3. Use 2 Master Glide Trauma for users with heavy legs or with trauma patients.



4. Place some Nylon Master on top of the board to minimize the friction during the transfer.



5. Place the beds close together and synchronize the height. Remember to lock all relevant breaks.



6. At a calm pace, pull the user towards the receiving bed using a drawsheet.



7. Remove the Nylon Master by pulling the layer furthest from the user.

User Manual - Master Glide Trauma Board

Technical Specifications

Materials:

PE

The products meet the criteria in the following standards:

REACH

RoSH

DODD Frank

Article no. - Product Name:	Size:	Handles:	Max. User Weight:
MC 004-1030 Master Glide Trauma Board	L 100 x W 49 x H 0,5 cm	1	150 kg
MC 004-1032 Master Glide Traume Board - 160	L 160 x B 49 x H 0,5 cm	1	150 kg

WARNINGS



Never leave the user on the product when lying on the edge of the bed.



Never leave the product on the floor.



Max 20 cm between the support surfaces during transfer with Master Board.



Max 5 cm slope between the support surfaces.



Make sure that all relevant brakes are activated before transferring the user.



CE-certification:

This product is CE-approved and follows the harmonized standards. This means that the products are in compliance with the essential health and safety directives. The product(-s) comply with the standards in the Medical Device Regulation (EU) 2017/745 – class 1.

For more information, videos Quick Guide etc., visit www.master-care.dk



User Weight:

Recommended max. user weight: 150 kg.



Surface cleaning:

The surface of this product can be cleaned with a wet cloth and soap.



Detergents pH level:

This product tolerates detergents / disinfectants with a pH level between 5-9.



Disinfectants:

This product tolerates alcoholic disinfectants with a pH level between 5-9.



Maintenance:

Make visual inspections frequently and preferably before every use. Pay attention to torn fabric, seams and straps. Products with high friction or other deficiencies must be discarded.



User Manual:

Always read the User Manual thoroughly before using the product.



Disposal:

Always make sure to follow local guidelines when disposing or discarding the product.



Find us on:

